

Amendments to the Specification:

Paragraph 24 is amended as follows:

[0024] The surgeon could use the exiting product as a means to gage if he is advancing too fast. A slow cutting motion will have some fat coming back up, and a fast expulsion through the guide rod end would indicate a too high canal pressure condition, indicating a need to slow the cutter advancement.

Figure 1 shows a cylindrical reamer 100 on a flexible shaft 200 with a power drill 300 to cut away bone.